

REPORT OF THE COUNCIL ON SCIENCE AND PUBLIC HEALTH

CSAPH Report 02-A-23

Subject: Improving Research Standards, Approval Processes, and Post-Market Surveillance Standards for Medical Devices

Presented by: Noel Deep, MD, Chair

Referred to: Reference Committee E

1 INTRODUCTION

2
3 Resolution 523-A-22, “Improving Research Standards, Approval Processes, and Post-Market
4 Surveillance Standards for Medical Devices” was referred by the House of Delegates (HOD). This
5 report serves as the Council on Science and Public Health’s (CSAPH) findings and
6 recommendations regarding medical device regulation.
7

8 METHODS

9
10 English language articles were selected from searches of PubMed and Google Scholar using the
11 search terms “medical device AND 510(k)” and “medical device AND post-market surveillance”.
12 Additional articles were identified by manual review of the reference lists of pertinent publications.
13 Web sites managed by government agencies and applicable organizations were also reviewed for
14 relevant information.
15

16 BACKGROUND

17
18 In the context of regulatory oversight by the Food and Drug Administration (FDA), a medical
19 device has a broad definition. According to the Food, Drug and Cosmetic Act:

20
21 a device is: an instrument, apparatus, implement, machine, contrivance, implant, in vitro
22 reagent, or other similar or related article, including a component part, or accessory which is:
23 [...]

24 (B) intended for use in the diagnosis of disease or other conditions, or in the cure, mitigation,
25 treatment, or prevention of disease, in man or other animals, or

26 (C) intended to affect the structure or any function of the body of man or other animals, and
27 which does not achieve its primary intended purposes through chemical action within or on the
28 body of man or other animals and which does not achieve its primary intended purposes
29 through chemical action within or on the body of man or other animals and which is not
30 dependent upon being metabolized for the achievement of its primary intended purposes.
31

32 As such, the breadth of items captured within this regulatory framework is expansive: ranging from
33 tongue depressors and eyeglasses to x-ray machines and hip replacements. In addition to physical
34 objects used as medical devices, software and algorithms are also captured within this definition.

35 As such, the FDA classifies software into two broad categories: software *in* a medical device and

© 2023 American Medical Association. All rights reserved.

1 software *as* a medical device (SaMD). CSAPH recognizes that software, particularly SaMD, is
2 rapidly becoming a large part of medical care and may warrant further examination beyond the
3 findings and recommendations of this report, which are intended to be generalizable to all medical
4 devices.

5 6 THE 510(K) REGULATORY PATHWAY 7

8 When applying for a new medical device, the device is first evaluated for risk category: I (lowest
9 risk), II (medium risk) or III (highest risk). Risk category is determined by a variety of factors, such
10 as by comparing the device to a similar, known, device. If a device is found to be like a device
11 already approved by the FDA, it may be classified as low (class I) or medium (class II) risk.
12 Examples of devices commonly found to be class I include electric toothbrushes, tongue
13 depressors, bandages, hospital beds, and non-electric wheelchairs. Examples of devices commonly
14 found to be class II include catheters, pregnancy test kits, syringes, contact lenses, and surgical
15 gloves. Examples of devices commonly found to be class III include breast implants, pacemakers,
16 defibrillators, and cochlear implants. Approximately 1% of all new medical device applications
17 from 2003 to 2017 were evaluated as high risk (class III).^{1,2}

18
19 If a medical device is found to be class I they are typically exempt from normal testing. If deemed
20 a class II risk, manufacturers may submit a 510(k) application as pre-market notification (PMN) to
21 the FDA. Class II risk devices are subjected to an equivalence evaluation comparing this product to
22 one currently on the market through these 510(k) processes. 510(k) applications are processed
23 within 90 days and once approved, the device is eligible for market. By contrast, class III devices
24 must undergo pre-market approval (PMA) which requires two large clinical trials. According to a
25 2010 industry survey, pursuing pre-market approval in the United States takes on average 54
26 months to complete compared to 11 months in European countries.³

27
28 Medical device market approval differs from drug approval in a few critical ways, which may help
29 illustrate why the 510(k) pathway is so desirable for medical device manufacturers. Table 1 in the
30 appendix of this report highlights some of these differences. Clinical trial design for medical
31 devices can be extremely difficult, and in some cases unethical. For example, a placebo control for
32 a medical device could require a high-risk sham surgery. As such, subjecting all new medical
33 devices to undergo clinical trials may substantially hinder innovation, particularly from physicians
34 seeking small tweaks or customizations to products they use routinely.

35
36 But on the other hand, if a medical device does cause harm to a patient, one cannot simply
37 discontinue having an implanted device without significant intervention unlike if they were
38 experiencing adverse events to a new medication that could be quickly stopped. As such, the
39 510(k) pathway has been subject to intense public scrutiny, both in the media and by elected
40 officials.⁴ Many recalls of medical devices are voluntarily initiated by the manufacturer due to
41 liability concerns or public perception decreasing sales rather than by official FDA action.

42
43 The FDA has recently begun piloting a new program within the 510(k) framework, called the
44 Safety and Performance Based Pathway. This pathway provides an alternative to the current
45 equivalence evaluation for a small subset of devices that are highly studied and well-known. In the
46 Safety and Performance Based Pathway, the FDA sets forth explicit benchmarks that medical
47 devices must satisfy to demonstrate safety and efficacy to gain 510(k) approval.⁵ For example, if a
48 resorbable surgical sutures manufacturer wished to market a new design, the FDA has guidance for
49 the appropriate diameter, needle attachment, tensile strength, sterilization, shelf life and resorption
50 profile for new suture designs to meet to receive 510(k) classification.⁶ This pathway provides

1 added safety and efficacy requirements to this moderate risk class. However, participation in the
2 Safety and Performance Based Pathway is currently optional.

3 4 DEVICE EQUIVALENCE

5
6 To be eligible for the 510(k) approval, a manufacturer must first establish that their device is
7 “substantially equivalent” to a previously known, FDA-approved predicate device.⁷ For the
8 purposes of regulatory approval, the FDA considers both safety and functionality when
9 determining equivalence. First, they investigate whether the device is to be used for the same
10 primary purpose, and they then evaluate whether the device is expected to have a similar safety
11 profile. For example, if a device were to change its power source (such as hardwired vs.
12 rechargeable) with no other modifications, it would likely be deemed substantially equivalent.
13 Similarly, if the material of the device were to change to another material known to be safe to the
14 FDA, it is likely to be found substantially equivalent. A flowchart of the FDA decision making
15 process has been included in the Appendix of this report.

16
17 However, there is a flaw with the approach of substantial equivalence. If a device is found to be
18 unsafe after receiving market approval and then subjected to a recall, any *subsequent* devices which
19 used the original, now-unsafe, device as their predicate are not subjected to any increased scrutiny
20 or recalls. Recent analysis found that between the period of 2017 and 2021, the FDA initiated
21 recalls of 156 devices using their highest risk categorization – devices with a reasonable probability
22 to cause severe morbidity and mortality. Of those 156 devices recalled, 44.1 percent of them had
23 received 510(k) approval using substantial equivalence to a device that had also been the subject of
24 a recall.⁸ Further, 48.1 percent of devices recalled within the studied period have themselves been
25 used as the predicate for another device’s 510(k) approval. This post marketing safety information
26 and related devices draw significant attention to potential problems with the current 510(k)
27 approval process with a lack of criterion for granting approval for devices outside the most well-
28 studied and well-understood.

29 30 *Post-Market Surveillance*

31
32 It should be noted that the study described above only studied a cohort of devices which were the
33 subject of FDA-initiated recalls. There are likely a non-trivial number of devices that are still being
34 used as comparators for substantial equivalence that have been found to be unsafe and then
35 production halted or voluntarily recalled by the manufacturer. However, there is limited publicly
36 available information to monitor this risk. This scenario highlights the importance of rigorous post-
37 market surveillance for devices that have been approved using the 510(k) pathway.

38
39 Among the post-market surveillance activities required by the FDA is the reporting of adverse
40 events. Under Medical Device Reporting regulations (Title 21 Code of Federal Regulations part
41 803), manufacturers, importers, and device user facilities (such as a hospital, nursing home or
42 outpatient treatment facilities) are mandatory reporters to the FDA regarding serious device
43 malfunction, including death. Reports are made to the device manufacturer (if known) and the
44 FDA. Health care professionals, patients, and caregivers are able to report suspected adverse events
45 for medical devices using the FDA’s MedWatch portal
46 (<https://www.accessdata.fda.gov/scripts/medwatch/>).

47
48 Adverse events are viewable to health care professionals and the public using the FDA’s
49 Manufacturer and User Facility Device Experience (MAUDE) portal.⁹ However, a 2019 exposé
50 found that over 5 million incidents of reported adverse events were being kept from public view
51 using an internal “alternative summary reporting” repository rather than the publicly available

1 MAUDE database.¹⁰ Not only did this practice prevent physicians and patients from knowing the
2 real risks of currently approved medical devices, it also prevented manufacturers of new devices
3 from knowing the risk profile of substantially similar predicate devices they were using for 510(k)
4 approval. The FDA has stated that it has since abandoned this practice of internal incident report
5 storage.¹¹

6 7 *Health Equity Considerations*

8
9 It should also be noted that implicit in the 510(k) substantial equivalence method of approval is that
10 it tends to maintain the status quo. For example, most, if not all, pulse oximeters currently used in
11 the United States are approved via the 510(k) pathway.¹² Pulse oximeters estimate blood oxygen
12 saturation by shining light through the skin, typically on a fingertip or an ear lobe. Oxygenated
13 blood absorbs red light more efficiently than de-oxygenated blood, thus allowing for estimates of
14 oxygenation by simply measuring the amount of red light that passes through a tissue. However,
15 oxygenated blood is not the only thing that absorbs red light – melanin, melanosomes, and
16 melanocytes (ie, skin pigmentation), also absorb or scatter red light. A retrospective study found
17 that practitioners missed hypoxemia diagnoses in 11.7 percent of Black patients compared to 3.6
18 percent of white patients due to pulse oximetry overestimating blood oxygenation.¹³

19
20 In the context of the COVID-19 pandemic, that suggests that excluding other factors, Black
21 patients would be nearly 4-times less likely to receive oxygenation therapy such as a ventilator,
22 which could prevent progression to acute respiratory distress syndrome.¹⁴ As a result of these
23 findings, the Food and Drug Administration (FDA) released a safety communication indicating
24 oximeters may be less accurate in darker skin tones.¹⁵ The failure of pulse oximeters to accurately
25 measure oxygen saturation in all skin tones is a clear example of how inequity enters the health
26 care system from many sources and can cascade. For example, even if a provider wished to start a
27 patient on oxygenation therapy, Medicare reimbursement for supplemental oxygen therapy is only
28 approved if a patient has a blood oxygenation reading less than or equal to 89 percent, which is less
29 likely in Black patients if a pulse oximeter is used.¹⁶ In November 2022, the FDA hosted an
30 advisory committee meeting to discuss concerns of pulse oximeters and skin pigmentation. Dr.
31 Jesse Ehrenfeld, president-elect of the AMA, was a participant of this meeting and delivered
32 comments and recommendations on behalf of the AMA.

33
34 It is important to assess whether approving a new pulse oximeter design that reaches the same level
35 of performance as a predicate device is appropriate as our appreciation of inequity grows and some
36 categories of devices no longer match the values we wish to uphold.

37 38 *Off-Label Use of Medical Devices*

39
40 While the FDA has attempted to pilot programs, such as the Safety and Performance Based
41 Pathway, that would improve the balance of fostering innovation and patient safety, they may not
42 have the legislative authority or resources available to make these new programs mandatory.
43 Without authority to pursue reforms to medical device regulation, there are concerns that the FDA
44 may become more and more likely to begin regulating the practice of medicine to achieve similar
45 goals.

46
47 The FDA has the authority to ban medical devices if they present a substantial deception to patients
48 about the benefits or an unreasonable and substantial risk of injury. However, there are recent
49 concerns of misuse of the banning process. In 2020, the FDA published a rule banning the use of
50 electrical stimulation devices (ESD) for the treatment of self-injurious and/or aggressive
51 behavior.¹⁷ The FDA reported that the use of ESDs for this indication was unsafe and could lead to

1 significant physical and psychological harm. ESDs were still approved for other indications such as
2 smoking cessation.¹⁸ The approval of devices for specific indications while banning the same
3 device for others is, per AMA policy, the FDA regulating the practice of medicine. The AMA has
4 extensive policy and significant history defending the rights of physicians to practice medicine and
5 protect off-label prescribing of pharmaceuticals and devices.

6
7 Within the text of the FDA’s rule on banning ESDs for aggressive behavior, they cite the 510(k)
8 pathway as part of their justification for the banning of a specific indication, as they evaluate risk
9 of a device based on its intended function, not on all potential functionalities. For example, daily
10 wear vs. extended wear for gas permeable contact lenses are two separate risk categories.
11 Evaluation of “substantially similar” for the purposes of 510(k) approval includes analysis of
12 similar function. In 2021, the D.C. Circuit Court of Appeals overturned the ban, finding that the
13 FDA was in fact regulating the practice of medicine, per the holdings of *Judge Rotenberg*
14 *Educational Center v. United States Food and Drug Administration*.¹⁹

15 16 CONCLUSION

17
18 While the FDA has made strides in improving the 510(k) process for medical device approval, such
19 as through the Safety and Performance Based Pathway, recent data have shown serious safety
20 concerns. These safety concerns denote the need for the process to be re-examined to support the
21 purpose and benefits of accelerated pathways along with providing the FDA with the statutory
22 authority to address the larger, systemic issues without impeding on the practice of medicine.

23 24 RECOMMENDATIONS

25
26 The Council on Science and Public Health recommends the following be adopted, and the
27 remainder of the report be filed:

- 28
29 1. Our AMA believes that to support innovation while protecting patient safety, approval
30 pathways for medical devices should incorporate the following principles:
- 31 a. Evidence-based, measurable performance benchmarks, such as those used in the
32 Safety and Performance Based Pathway, should be used wherever possible for
33 classes of known, well-studied medical devices; and
 - 34 b. For a subset of higher risk devices receiving approval but have not completed
35 clinical trials, time-limited approvals may be appropriate, after which the
36 manufacturer may be required to provide post-market data to support full device
37 approval; and
 - 38 c. Medical devices with known safety concerns should not be usable as predicate
39 devices for the purposes of proving substantial equivalence. In the event safety
40 concerns of predicate devices arise after approval has been granted, additional due
41 diligence should be initiated as appropriate; and
 - 42 d. Approval for medical devices should include criteria for adequate performance in
43 racialized, minoritized, or otherwise historically excluded groups when feasible;
44 ~~and~~
 - 45 e. ~~Reports of adverse events for medical devices should always be available in a~~
46 ~~publicly accessible, searchable database such as the Manufacturer and User~~
47 ~~Facility Device Experience. (New HOD Policy)~~

- 1
 - 2
 - 3
2. That Policy H-120.988, "Patient Access to Treatments Prescribed by Their Physicians", supporting a physician's right to prescribe medical devices off-label, be reaffirmed. (Reaffirm Current HOD Policy)

APPENDIX

TABLE 1


Comparison of regulatory requirements for drugs, biologics, and devices

Modified from Congressional Research Service, “Medical Product Regulation: Drugs, Biologics, and Devices”, published September 29th, 2021. <https://sgp.fas.org/crs/misc/IF11083.pdf>.

	Drug	Biologic	Class II (Medium Risk) Device	Class III (High Risk) Device
Authorization Type	Approval	Licensure	Clearance	Approval
Submission to FDA	New Drug Application	Biologics License Application	510(k) notification	Pre-market approval
Clinical Trials?	Yes	Yes	No	Yes (few exceptions)
Evidence Required by FDA	Substantial evidence of effectiveness, adequate evidence of safety	Substantial evidence of effectiveness, adequate evidence of safety	Substantial equivalence to a known, approved device	Reasonable assurance that the device is safe and effective for its intended use(s)

FDA 510(k) Decision-Making Flowchart

Modified from Food and Drug Administration, “The 510(k) Program: Evaluating Substantial Equivalence in Premarket Notifications [510(k)]”, July 28, 2014. Accessed January 23rd, 2023.



SE = "Substantially Equivalent"
 NSE = "Not Substantially Equivalent"
 IFU = "Indications For Use"

REFERENCES

- ¹ Institute of Medicine of the National Academies. *Medical Devices and the Public's Health: The FDA 510(k) Clearance Process at 35 Years*. 2011.
- ² Dubin, Jonathan R., et al. *Risk of recall among medical devices undergoing US Food and Drug Administration 510 (k) clearance and premarket approval, 2008-2017*. JAMA Network Open 4.5 (2021): e217274-e217274.
- ³ Maak, Travis G., and James D. Wylie. *Medical device regulation: a comparison of the United States and the European Union*. Journal of the American Academy of Orthopaedic Surgeons 24, no. 8 (2016): 537-543.
- ⁴ New York Times Editorial Board. *80,000 deaths. 2 million injuries. It's time for a reckoning on medical devices*. New York Times. May 4, 2019. <https://www.nytimes.com/2019/05/04/opinion/sunday/medical-devices.html>.
- ⁵ Food and Drug Administration. *Safety and Performance Based Pathway*. September 20, 2019. <https://www.fda.gov/media/112691/download>. Accessed January 17, 2023.
- ⁶ Food and Drug Administration. *Surgical Sutures – Performance Criteria for Safety and Performance Based Pathway*. April 11, 2022. <https://www.fda.gov/media/157490/download>. Accessed January 17, 2023.
- ⁷ Food and Drug Administration, *The 510(k) Program: Evaluating Substantial Equivalence in Premarket Notifications [510(k)]*. July 28, 2014. <https://www.fda.gov/media/82395/download>. Accessed January 23, 2023.
- ⁸ Kadakia, Kushal T., et al. *Use of Recalled Devices in New Device Authorizations Under the US Food and Drug Administration's 510 (k) Pathway and Risk of Subsequent Recalls*. JAMA 329.2 (2023): 136-143.
- ⁹ Food and Drug Administration. *MAUDE - Manufacturer and User Facility Device Experience*. <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfmaude/search.cfm>.
- ¹⁰ Jewett, Christina. *Hidden FDA Reports Detail Harm Caused By Scores Of Medical Devices*. Kaiser Health News. March 7, 2019. <https://khn.org/news/hidden-fda-database-medical-device-injuries-malfunctions/>.
- ¹¹ Food and Drug Administration. *Statement on agency's efforts to increase transparency in medical device reporting*. June 21, 2019. <https://www.fda.gov/news-events/press-announcements/statement-agencys-efforts-increase-transparency-medical-device-reporting>
- ¹² McFarling, Usha Lee. *FDA panel asks for improvements in pulse oximeters*. STAT News. Nov 1, 2022. <https://www.statnews.com/2022/11/01/fda-panel-asks-for-improvements-in-pulse-oximeters/>.
- ¹³ Sjoding, Michael W., et al. *Racial bias in pulse oximetry measurement*. New England Journal of Medicine 383.25 (2020): 2477-2478.
- ¹⁴ Id.
- ¹⁵ Food and Drug Administration. *Pulse Oximeter Accuracy and Limitations: FDA Safety Communication*. Nov 7, 2022. <https://www.fda.gov/medical-devices/safety-communications/pulse-oximeter-accuracy-and-limitations-fda-safety-communication>.
- ¹⁶ ResMed. *Reimbursement fast facts: oxygen concentrators*. Accessed March 8, 2023. https://document.resmed.com/en-us/documents/articles/reimbursement_fast_facts_oxygen_concentrators_amer_eng.pdf.
- ¹⁷ Food and Drug Administration. *Banned Devices; Electrical Stimulation Devices for Self-Injurious or Aggressive Behavior*. March 6, 2020. <https://www.federalregister.gov/documents/2020/03/06/2020-04328/banned-devices-electrical-stimulation-devices-for-self-injurious-or-aggressive-behavior>.
- ¹⁸ Hale, Conor. *FDA bans electric shock devices for conditioning against aggressive behaviors*. Fierce Biotech. March 4, 2020. <https://www.fiercebiotech.com/medtech/fda-bans-electric-shock-devices-for-conditioning-against-aggressive-behaviors>.
- ¹⁹ Judge Rotenberg Educational Center v. United States Food and Drug Administration. No. 20-1087 (D.C. Cir. July 6, 2021). [https://www.cadc.uscourts.gov/internet/opinions.nsf/C32A7577ED02127D8525870A00555511/\\$file/20-1087-1905079.pdf](https://www.cadc.uscourts.gov/internet/opinions.nsf/C32A7577ED02127D8525870A00555511/$file/20-1087-1905079.pdf)